



Allegato CSAAR al Certificato di Accreditamento 0249ISP
Annex CSAAR to the Accreditation Certificate 0249ISP

RILASCIATO A
ISSUED TO **CMT S.n.c. di Bracaloni Alessandro & C.**

CMT Organismo di Ispezione

REVISIONE
REVISION **001** DATA DI REVISIONE
REVIEW DATE **28-02-2024** DATA DI SCADENZA
EXPIRY DATE **26-03-2028**

Ispezione di Tipo C nei seguenti settori:

Verificazione periodica di:

- strumenti per pesare a funzionamento non automatico – NAWI – classe I – con portata massima fino a 1 kg;
- strumenti per pesare a funzionamento non automatico – NAWI – classe II – con portata massima fino a 40 kg;
- strumenti per pesare a funzionamento non automatico – NAWI – classi III e IIII – con portata massima fino a 300 kg;

ai sensi del Decreto 21 aprile 2017, n. 93.

Inspection of Type C in the following sectors:

Subsequent verification of:

- non-automatic weighing instruments – NAWI – class I – with maximum capacity up to 1 kg;
- non-automatic weighing instruments – NAWI – class II – with maximum capacity up to 40 kg;
- non-automatic weighing instruments – NAWI – classes III and IIII – with maximum capacity up to 300 kg;

in accordance with the Decree 21 april 2017, n. 93.